



SINCE 1978

OPTICS TECHNOLOGY

Manufacturer & Exporters of : LAB TESTING INSTRUMENTS



An ISO 9001:2015 CERTIFIED IEC 61010-1:2010
An ISO 10002:2018 CERTIFIED BS EN 12469:2000
An ISO 13485:2016 CERTIFIED ISO 29463-5:2022

ANTIBIOTIC ZONE READER CE

DIGITAL MODEL



OT-AZR-165/D

DIGITAL MODEL

Most suitable to measure the diameter of the inhibited zone in microbiology, botany, zoology, bacteriology and pathology labs. Unique in design and measures the diameter of the inhibited zone to 0.1 mm within the range of 0 to 35 mm. Light from a source in the base passes up through the transparent and Semi transparent portion of the agar-agar, then to a reflecting mirror supported by an arm above the unit. The mirror reflects the light to a glass prism mounted at the front of the unit and the magnified image of the zone of inhibitions can be clearly seen on the prism. Supplied complete with accessories like prism, petri dish. Suitable for operation on 230 V, 50 Hz, single phase AC supply.

Digital Model:

Range : 0.0 to - 39.9 mm.

Optional:

1. S.S. Calibrated Coins

2/3, DSIDC, (Behind Fun Cinema) NSP, Pitampura, Delhi-110034 (INDIA)

☎ 9811238150, 9811238151, 9811236960

E-mail: info@opticstechnology.in

opticstechnology@gmail.com

Website : www.opticstechnology.in