



# OPTICS TECHNOLOGY



SINCE 1978

Manufacturer & Exporters of : LAB TESTING INSTRUMENTS



An ISO 9001:2015 CERTIFIED IEC 61010-1:2010  
An ISO 10002:2018 CERTIFIED BS EN 12469:2000  
An ISO 13485:2016 CERTIFIED ISO 29463-5:2022

## BACTERIOLOGICAL INCUBATOR CE

(WITH TOUCH SCREEN DISPLAY & DATA LOGGING)



OT-BI-56 TS/SS



TOUCH SCREEN MODEL



OT-BI-56 TS/SS

Pen Drive

**Double walled**, Inner chamber made of Stainless Steel 304. Outside Mild Steel painted in epoxy powder coating / stainless steel 304 with wire mesh adjustable shelves, two indicators and on /off switch to work on 220/230 volts A.C. Heating elements are placed in ribs at the bottom and sides. Temperature range Ambient Temp. **+5°C to 70°C**. The door has synthetic rubber gasket. With one number Air circulating Fan. Specially Designed to maintain 37°C temp.

- Touch Screen Display Menu driven settings for Controlling parameters Setting / RTC accurate Date Time Clock.
- Pen Drive for saving of data for printing through PC communication.

Model No.	Size Inside Chamber W x D x H	Ltrs. (App.)	No. of Trays
OT-BI-56/28-MS/SS/TS	300 x 300 x 300 mm (12"x12"x12")	28	1
OT-BI-56/45-MS/SS/TS	355 x 355 x 355 mm (14"x14"x14")	45	1
OT-BI-56/96-MS/SS/TS	450 x 450 x 450 mm (18"x18"x18")	96	2
OT-BI-56/128-MS/SS/TS	450 x 450 x 600 mm (18"x18"x24")	128	2
OT-BI-56/227-MS/SS/TS	600 x 600 x 600 mm (24"x24"x24")	227	2
OT-BI-56/255-MS/SS/TS	450 x 600 x 900 mm (24"x18"x36")	255	3
OT-BI-56/340-MS/SS/TS	600 x 600 x 900 mm (24"x24"x36")	340	3
OT-BI-56/453-MS/SS/TS	600 x 600 x 1200 mm (24"x24"x48")	453	4

OR ANY NON STANDARD SIZE

### Optional :

1. Special heavy duty air circulating fan system.
2. Online transfer of data without any cable/wire Interference
3. 21 CFR software (for non-editable data)

2/3, DSIDC, (Behind Fun Cinema) NSP, Pitampura, Delhi-110034 (INDIA)

☎ 9811238150, 9811238151, 9811236960

E-mail: [info@opticstechnology.in](mailto:info@opticstechnology.in),

[opticstechnology@gmail.com](mailto:opticstechnology@gmail.com),

Website : [www.opticstechnology.in](http://www.opticstechnology.in)